

PHILADELPHIA, PA 19103-6996

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/666,290	09/19/2003	Arjunan Ganesh	08321-0099 US1	9071
23973 . 7:	590 08/11/2005		EXAM	INER
DRINKER BI	IDDLE & REATH	ALI, SHUMAYA B		
ATTN: INTEL	·			
ONE LOGAN SQUARE		•	ART UNIT	PAPER NUMBER
18TH AND CHERRY STREETS		3743		

DATE MAILED: 08/11/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

	Application No.	Applicant(s)			
	10/666,290	GANESH ET AL.			
Office Action Summary	Examiner	Art Unit			
	Shumaya B. Ali	3743			
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address			
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).					
Status					
1) Responsive to communication(s) filed on <u>19 September 2003</u> .					
2a) This action is FINAL . 2b) ⊠ This	This action is FINAL . 2b)⊠ This action is non-final.				
3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is					
closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.					
Disposition of Claims					
 4) Claim(s) 1-16 is/are pending in the application. 4a) Of the above claim(s) is/are withdraw 5) Claim(s) is/are allowed. 6) Claim(s) 1-16 is/are rejected. 7) Claim(s) 8,10 and 15 is/are objected to. 8) Claim(s) are subject to restriction and/or 					
Application Papers	·				
9)☐ The specification is objected to by the Examiner 10)☒ The drawing(s) filed on 19 September 2003 is/a Applicant may not request that any objection to the of Replacement drawing sheet(s) including the correction 11)☐ The oath or declaration is objected to by the Examiner	re: a)⊠ accepted or b)□ object drawing(s) be held in abeyance. See on is required if the drawing(s) is obj	e 37 CFR 1.85(a). ected to. See 37 CFR 1.121(d).			
Priority under 35 U.S.C. § 119					
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 					
Attackment(s)					
Attachment(s) Notice of References Cited (PTO-892) Notice of Draftsperson's Patent Drawing Review (PTO-948) Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) Paper No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail Da 5) Notice of Informal P 6) Other: detailed action	ate atent Application (PTO-152)			

Application/Control Number: 10/666,290 Page 2

Art Unit: 3743

DETAILED ACTION

Claim Rejections - 35 USC § 112

1. The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

2. Claims 8,10,15 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter, which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. Limitations cited in claims 8,10, and 15 are inconsistent or may not be found in the original disclosure.

Claim Rejections - 35 USC § 103

- 3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 1-7,9,11-12,14, and 16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Christopher US Patent application publication US 2002/0108610A1

4. As to claim 1, Christopher anticipates claimed limitation in figs. 1 and 2: a body (25) which is depicted as a tube which inherently have a distallend (insertable in the mouth) and a proximal end (situates

Application/Control Number: 10/666,290 Page 3

Art Unit: 3743

outside of the mouth), since the length of the tube can be controlled by an operator and also predetermined depending on the surgical procedures, the tube can be sized and shaped to be disposed within the pharynx above the epiglottis; at least one channel (lumen of the body) forming at least one airway in the device body extending between the proximal end and the distal end of the device body; at least one first conduit (130) in the device body for conveying an inhalant gas to the patient that extends from the proximal end to the distal end of the device body; at least one second conduit (it would have been obvious to include a conduit for suction secretion from the patient's mouth and airway, see page 3, 0039) for suctioning that extends from the proximal end to the distal end of the device body; and at least one third conduit (140) for sampling gas exhaled by the patient that extends from the proximal end of the device body and terminates at a position in the channel.

- 5. As to claim 2, Christopher discloses the oropharyngeal device according to claim 1 wherein at least one of the at least one first, second, and third conduits is disposed within the device body (see fig.1 and 2).
- 6. As to claims 3-6, the limitations are not considered to be critical since the where the conduits will be situated long the length of the body may vary between patients and procedures. Therefore, meeting one of the limitations cited in claims 3-6 would considered to meet all claimed limitation. Christopher discloses "with the guide or formed as a conduit" (claim 4) page 3, 0034.
- 7. As to claim 7, Christopher discloses the oropharyngeal device according to claim 1 wherein the at least one channel has a U-shaped cross section (channel where conduits 130 and 140 are situated are considered to have U-shaped cross second).
- 8. As to claim 9, Christopher discloses the oropharyngeal device according to claim 1 wherein the device is rigid and functions as a bite block (pg.2 0031, lines 1-4).

Application/Control Number: 10/666,290 Page 4

Art Unit: 3743

9. As to claim 11, Christopher discloses the oropharyngeal device according to claim 1 wherein the first, second, and third conduits are coupled to connectors (pg. 3, 0040 line 8) at the proximal end of the device for connecting to an inhalant gas source, a suctioning device, and a gas sampling device, respectively (fig.1 seems to depict all conduits 130, 140 have connecting means at the proximal end of the conduits that allows connecting of the conduit to air supplies)

- 10. As to claim 12, Christopher discloses the oropharyngeal device according to claim 1 further comprising at least one flexible extension (conduit are considered extension conduit due to having connector means) conduit coupled to at least one of the first, second, and third conduits at the proximal end of the device body (all conduits are considered flexible since they are insertable through a bent tube).
- 11. As to claim 14, Christopher does not disclose "right angled connector". Since connectors comes in various sizes and shapes it would have been obvious to one of ordinary skills in the art to substitute one type of connectors with another for the purposes of meeting their needs.

Claim 13 is rejected under 35 U.S.C. 103(a) as being unpatentable over Christopher US Patent application publication US 2002/0108610A1 in view of Kinner et al. US Patent 3,756,244

12. As to claim 13, Christopher does not disclose, "flange", however Kinner et al. teach flanges (fig.2, 16) at one end of a breathing aid. Therefore it would have been obvious to one of ordinary skills in the art at the time the invention was made to modify the device of Christopher in view of Kinner et al. in order to include a flange for the purposes of preventing the device from further entering the mouth or extending deeper into the pharynx and avoiding choking (col.1 lines 14-17).

Claim Objections

13. Claims 8,10, and 15 are objected to under 37 CFR 1.75(c), as being of improper dependent form for failing to further limit the subject matter of a previous claim. Applicant is required to cancel the claim(s), or amend the claim(s) to place the claim(s) in proper dependent form, or rewrite the claim(s) in independent form. Subject matter of the claims is vague and unclear in the disclosure.

Conclusion

14. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

US Patent 2127215 discloses flange, conduit, U-shaped tube, and connector; US Patent 4231365

discloses connector, conduit, and flange, and US Patent 6568388 discloses connector and conduit.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Shumaya B. Ali whose telephone number is 571-272-6088. The examiner can normally be reached on M-F 8:30-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Henry Bennett can be reached on 571-272-4791. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Art Unit: 3743

Shumaya B Ali Examiner Art Unit 3743